



America

CERTIFICATE

No. Z1 13 10 14544 102

Holder of Certificate: **Fluke Corporation**

6920 Seaway Boulevard
Everett, WA 98203-9090
USA

Production Facility(ies):

14544, 56588

Certification Mark:



Product:

**Multimeters
(Remote Display Handheld Digital Multimeter)**

Model(s):

233

Parameters:

Rated Input Voltage:	4.5 V DC Meter (3 x 1.5V AA batteries IEC LR6) 3.0 V DC Display (2 x 1.5V AA batteries IEC LR6)
Measuring Ratings:	600V CAT IV / 1000 V CAT III
Protection Class:	II
IP Protection:	IPX0
Temperature Ambient:	-10°C to +50°C

Tested according to:

EN 61010-1:2010
EN 61010-2-030:2010

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

Test report no.:

PI1305384105-000

Date, 2013-10-10

Page 1 of 1

